

Version 9 Last updated 30 July 2018

# ab100647 TGF beta 1 Human ELISA Kit

For the quantitative measurement of human TGF beta 1 in serum, plasma, and cell culture supernatants.

This product is for research use only and is not intended for diagnostic use.

## Table of Contents

|                                     |    |
|-------------------------------------|----|
| 1. Overview                         | 1  |
| 2. Protocol Summary                 | 2  |
| 3. Precautions                      | 3  |
| 4. Storage and Stability            | 3  |
| 5. Limitations                      | 4  |
| 6. Materials Supplied               | 4  |
| 7. Materials Required, Not Supplied | 5  |
| 8. Technical Hints                  | 6  |
| 9. Reagent Preparation              | 7  |
| 10. Standard Preparation            | 8  |
| 11. Sample Preparation              | 10 |
| 12. Plate Preparation               | 11 |
| 13. Assay Procedure                 | 12 |
| 14. Typical Data                    | 13 |
| 15. Typical Sample Values           | 15 |
| 16. Assay Specificity               | 16 |
| 17. Species Reactivity              | 16 |
| 18. Troubleshooting                 | 17 |
| 19. Notes                           | 18 |

# 1. Overview

Abcam's TGF beta 1 Human ELISA (Enzyme-Linked Immunosorbent Assay) kit is an in vitro enzyme-linked immunosorbent assay for the quantitative measurement of Human TGF beta 1 in serum, plasma and cell culture supernatants.

This assay employs an antibody specific for Human TGF beta 1 coated on a 96-well plate. Standards and samples are pipetted into the wells and TGF beta 1 present in a sample is bound to the wells by the immobilized antibody. The wells are washed and biotinylated anti-Human TGF beta 1 antibody is added. After washing away unbound biotinylated antibody, HRP-conjugated streptavidin is pipetted to the wells. The wells are again washed, a TMB substrate solution is added to the wells and color develops in proportion to the amount of TGF beta 1 bound. The Stop Solution changes the color from blue to yellow, and the intensity of the color is measured at 450 nm.

## 2. Protocol Summary

Prepare all reagents, samples, and standards as instructed



Add standard or sample to appropriate wells.

Incubate at room temperature.



Wash and add prepared biotin antibody to each well. Incubate at room temperature.



Wash and add prepared Streptavidin Solution. Incubate at room temperature.



Add TMB One-Step Development Solution to each well. Incubate at room temperature



Add Stop Solution to each well. Read at 450 nm immediately.

### 3. Precautions

**Please read these instructions carefully prior to beginning the assay.**

- All kit components have been formulated and quality control tested to function successfully as a kit.
- We understand that, occasionally, experimental protocols might need to be modified to meet unique experimental circumstances. However, we cannot guarantee the performance of the product outside the conditions detailed in this protocol booklet.
- Reagents should be treated as possible mutagens and should be handled with care and disposed of properly. Please review the Safety Datasheet (SDS) provided with the product for information on the specific components.
- Observe good laboratory practices. Gloves, lab coat, and protective eyewear should always be worn. Never pipet by mouth. Do not eat, drink or smoke in the laboratory areas.
- All biological materials should be treated as potentially hazardous and handled as such. They should be disposed of in accordance with established safety procedures.

### 4. Storage and Stability

**Store kit at -20°C immediately upon receipt. Kit has a storage time of 1 year from receipt, providing components have not been reconstituted.**

Refer to list of materials supplied for storage conditions of individual components. Observe the storage conditions for individual prepared components in the Materials Supplied section.

## 5. Limitations

- Assay kit intended for research use only. Not for use in diagnostic procedures.
- Do not mix or substitute reagents or materials from other kit lots or vendors. Kits are QC tested as a set of components and performance cannot be guaranteed if utilized separately or substituted.

## 6. Materials Supplied

| Item                                  | Quantity | Storage Condition |
|---------------------------------------|----------|-------------------|
| TGF beta 1 Microplate (12 x 8 wells)  | 96 wells | -20°C             |
| 20X Wash Buffer Concentrate           | 25 ml    | -20°C             |
| 5X Assay Diluent                      | 15 ml    | -20°C             |
| Biotinylated anti-Human TGF beta 1    | 2 vials  | -20°C             |
| Recombinant Human TGF beta 1 Standard | 2 vials  | -20°C             |
| 500X HRP-Streptavidin Concentrate     | 200 µl   | -20°C             |
| TMB One-Step Substrate Reagent        | 12 ml    | -20°C             |
| Stop Solution                         | 8 ml     | -20°C             |

## 7. Materials Required, Not Supplied

These materials are not included in the kit, but will be required to successfully perform this assay:

- 1 Microplate reader capable of measuring absorbance at 450 nm.
- Precision pipettes to deliver 2  $\mu$ L to 1 mL volumes.
- Adjustable 1-25 mL pipettes for reagent preparation.
- 100 mL and 1 liter graduated cylinders.
- Absorbent paper.
- Distilled or deionized water.
- Log-log graph paper or computer and software for ELISA data analysis.
- Tubes to prepare standard or sample dilutions.
  
- Reagents required for Sample activation:
  - o **1 N HCl (100 ml)** – Slowly add 8.33 mL of 12 N HCl into 91.67 ml deionized water. Mix bottle.
  - o **1.2 N NaOH/0.5 M HEPES (100 ml)** - Slowly add 12 ml of 10 N NaOH into 75 mL deionized water. Mix bottle. Add 11.9 g HEPES. Mix through. Bring final volume to 100 mL with deionized water.

## 8. Technical Hints

- This kit is sold based on number of tests. A 'test' simply refers to a single assay well. The number of wells that contain sample, control or standard will vary by product. Review the protocol completely to confirm this kit meets your requirements. Please contact our Technical Support staff with any questions.
- Selected components in this kit are supplied in surplus amount to account for additional dilutions, evaporation, or instrumentation settings where higher volumes are required. They should be disposed of in accordance with established safety procedures.
- Ensure plates are properly sealed or covered during incubation steps.
- When preparing your standards, it is critical to briefly centrifuge the vial first. The powder may adhere to the cap and not be included in the standard solution resulting in an incorrect concentration. Be sure to dissolve the powder thoroughly when reconstituting. After adding Assay Diluent to the vial, we recommend inverting the tube a few times, then flick the tube a few times, and centrifuge briefly; repeat this procedure 3-4 times. This is an effective technique for thorough mixing of the standard without using excessive mechanical force.
- Do not vortex the standard during reconstitution, as this will destabilize the protein.
- Once your standard has been reconstituted, it should be used right away or else frozen for later use.
- Keep the standard dilutions on ice while during preparation, but the ELISA procedure should be done at room temperature.
- Be sure to discard the working standard dilutions after use – they do not store well.
- Samples generating values higher than the highest standard should be further diluted in the appropriate sample dilution buffers.
- Avoid foaming or bubbles when mixing or reconstituting components.
- Avoid cross contamination of samples or reagents by changing tips between sample, standard and reagent additions.
- Ensure plates are properly sealed or covered during incubation steps.
- Complete removal of all solutions and buffers during wash steps.



- Make sure the microplate reader is switched on before starting the experiment.

## 9. Reagent Preparation

- Equilibrate all reagents to room temperature (18-25°C) prior to use. The kit contains enough reagents for 96 wells.
- Prepare only as much reagent as is needed on the day of the experiment.

### 9.1 1X Assay Diluent

5X Assay Diluent should be diluted 5-fold with deionized or distilled water before use.

### 9.2 1X Wash Solution

If the 20X Wash Concentrate contains visible crystals, equilibrate to room temperature and mix gently until dissolved. Dilute 20 ml of 20X Wash Solution Concentrate into deionized or distilled water to yield 400 ml of 1X Wash Solution.

### 9.3 1X Biotinylated TGF beta 1 Detection Antibody

Briefly spin the Biotinylated anti-Human TGF beta 1 vial before use. Add 100 µL of 1X Assay Diluent into the vial to prepare a detection antibody concentrate. Pipette up and down to mix gently (the concentrate can either be stored at 4°C for 5 days or aliquoted and frozen at -20°C for 2 months). The detection antibody concentrate must be diluted 80-fold with 1X Assay Diluent prior to use in the Assay Procedure.

### 9.4 1X HRP-Streptavidin Solution

Briefly spin the 500X HRP-Streptavidin concentrate vial before use. HRP-Streptavidin concentrate must be diluted 500-fold with 1X Assay Diluent prior to use in the Assay Procedure.

## 10. Standard Preparation

- Prepare serially diluted standards immediately prior to use. Always prepare a fresh set of standards for every use.
- Standard (recombinant protein) should be stored at -20°C or -80°C (recommended at -80°C) after reconstitution.
- The following section describes the preparation of a standard curve for duplicate measurements (recommended).

**10.1** Briefly spin the vial of TGF beta 1 Standard. Prepare a 125 ng/mL TGF beta 1 **Stock Standard** by adding 400 µL Assay Diluent.

**10.2** Ensure the powder is thoroughly dissolved by gentle mixing.

**10.3** Label tubes #1 – 8.

**10.4** Prepare **Standard #1** by adding 20 µL of the 125 ng/mL Stock Standard to 605 µL Assay Diluent into tube 1#. Mix thoroughly and gently.

**10.5** Pipette 300 µL Assay Diluent into each tube.

**10.6** Prepare **Standard #2** by transferring 200 µL from tube #1 to #2, mix thoroughly.

**10.7** Prepare **Standard #3** by transferring 200 µL from tube #2 to #3, mix thoroughly.

**10.8** Using the table below as a guide, prepare subsequent serial dilutions. Standard #8 contains no protein and is the Blank control.

| Standard # | Volume to dilute (μL) | Diluent (μL) | Total Volume (μL) | Starting Conc. (pg/mL) | Final Conc. (pg/mL) |
|------------|-----------------------|--------------|-------------------|------------------------|---------------------|
| 1          | 20                    | 605          | 625               | 125ng/ml               | 4000                |
| 2          | 200                   | 300          | 500               | 4000                   | 1600                |
| 3          | 200                   | 300          | 500               | 1600                   | 640                 |
| 4          | 200                   | 300          | 500               | 640                    | 256                 |
| 5          | 200                   | 300          | 500               | 256                    | 102.4               |
| 6          | 200                   | 300          | 500               | 102.4                  | 40.96               |
| 7          | 200                   | 300          | 500               | 40.96                  | 16.38               |
| 8 (Blank)  | 0                     | 300          | 300               |                        |                     |

## 11. Sample Preparation

### 11.1 TGF beta 1 Sample Activation Procedure

To activate latent TGF beta 1 to the immunoreactive form, follow the activation procedure outlined below. Assay samples after neutralization (pH 7.0 – 7.6). Use polypropylene test tubes.

**Δ Note: Do not activate the kit standards. The kit standards contain active rhTGF beta 1.**

### 11.2 Cell Culture Supernates

Add 0.1 mL 1 N HCL into 0.5 mL cell culture supernate. Mix tube thoroughly. Incubate for 10 minutes at room temperature. Neutralize the acidified sample by adding 0.1 mL 1.2 N NaOH/0.5 M HEPES (pH 7.0 – 7.6). Mix tubes thoroughly. Assay immediately. The activated sample may be diluted with 1X Assay Diluent. The concentration read off the standard curve must be multiplied by the dilution factor.

### 11.3 Serum/Plasma

Add 0.05 mL 1 N HCl to 0.1 mL serum. Mix tube thoroughly. Incubate for 10 minutes at room temperature. Neutralize the acidified sample by adding 0.05 mL 1.2 N NaOH/0.5 M HEPES. Mix tube thoroughly. Assay immediately. The activated sample may be diluted with 1X Assay Diluent. The concentration read off the standard curve must be multiplied by the dilution factor.

## 12. Plate Preparation


- The 96 well plate strips included with this kit are supplied ready to use. It is not necessary to rinse the plate prior to adding reagents.
- Unused well strips should be returned to the plate packet and stored at 4°C.
- For each assay performed, a minimum of 2 wells must be used as blanks, omitting primary antibody from well additions.
- For statistical reasons, we recommend each sample should be assayed with a minimum of two replicates (duplicates).
- Well effects have not been observed with this assay.

## 13. Assay Procedure

- Equilibrate all materials and prepared reagents to room temperature prior to use.
  - We recommend that you assay all standards, controls and samples in duplicate.
- 13.1** Add 100 µl of each standard (see Standard Preparation section 10) and sample into appropriate wells. Cover well and incubate for 2.5 hours at room temperature or overnight at 4°C with gentle shaking.
  - 13.2** Discard the solution and wash 4 times with 1X Wash Solution. Wash by filling each well with 1X Wash Solution (300 µl) using a multi-channel Pipette or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining 1X Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
  - 13.3** Add 100 µL of 1X Biotinylated TGF beta 1 Detection Antibody (Reagent Preparation section 9) to each well. Incubate for 1 hour at room temperature with gentle shaking.
  - 13.4** Discard the solution. Repeat the wash as in step 13.2.
  - 13.5** Add 100 µL of 1X HRP-Streptavidin solution (see Reagent Preparation section 9) to each well. Incubate for 45 minutes at room temperature with gentle shaking.
  - 13.6** Discard the solution. Repeat the wash as in step 13.2.
  - 13.7** Add 100 µl of TMB One-Step Substrate Reagent to each well. Incubate for 30 minutes at room temperature in the dark with gentle shaking.
  - 13.8** Add 50 µl of Stop Solution to each well. Read at 450 nm immediately.
  - 13.9** Analyze the data as described below.
    - 13.9.1** Calculate the mean absorbance for each set of duplicate standards, controls and samples, and subtract the average Blank absorbance value.
    - 13.9.2** Plot the standard curve on log-log graph paper, with standard concentration on the x-axis and absorbance on the y-axis. Draw the best-fit straight line through the standard points.
    - 13.9.3** Determine the unknown sample concentration from the Standard Curve and multiply the value by the dilution factor.

## 14. Typical Data

Typical standard curve – data provided for demonstration purposes only. A new standard curve must be generated for each assay performed.



## 15. Typical Sample Values

### SENSITIVITY –

The minimum detectable dose (MDD) of TGF beta 1 is typically 18 pg/mL.

### PRECISION –

|        | Intra-Assay | Inter-Assay |
|--------|-------------|-------------|
| CV (%) | <10%        | <12%        |

### RECOVERY –

Recovery was determined by spiking various levels of recombinant Human TGF beta 1 into Human serum, plasma and cell culture media. Mean recoveries are as follows:

| Sample Type        | Average % Recovery | Range (%) |
|--------------------|--------------------|-----------|
| Serum              | 95.80              | 83-103    |
| Plasma             | 83.71              | 75-94     |
| Cell Culture Media | 121.7              | 113-127   |



## Linearity of Dilution

| Serum Dilution | Average % Expected Value | Range (%) |
|----------------|--------------------------|-----------|
| 1:2            | 124.5                    | 118-131   |
| 1:4            | 98.08                    | 77-120    |

| Plasma Dilution | Average % Expected Value | Range (%) |
|-----------------|--------------------------|-----------|
| 1:2             | 121.2                    | 116-125   |
| 1:4             | 133.5                    | 129-141   |

| Cell Culture Media Dilution | Average % Expected Value | Range (%) |
|-----------------------------|--------------------------|-----------|
| 1:2                         | 103.9                    | 88-119    |
| 1:4                         | 105.7                    | 97-114    |

## 16. Assay Specificity

The antibodies used within this ELISA kit detect human TGF beta 1.

Cross Reactivity: This ELISA kit shows no cross-reactivity with any of the cytokines tested; ANG, CD23, Eotaxin, GCSF, GM-CSF, GRO- $\alpha$ , GRO- $\beta$ , GRO- $\gamma$ , I-309, IFN- $\gamma$ , IL-1 $\alpha$ , IL-1 $\beta$ , IL-3, IL-4, IL-5, IL-6, IL-7, IL-8, IL-10, IL-12 (p40), IL-12 (p70), IL-15, IL-16, IP-10, MCP-1, MCP-2, MCP-3, MCP-4, MCSF, MIG, MIP-1 $\alpha$ , MIP-1 $\beta$ , NAP-2, PDGF, PF-4, PARC, SCF, SDF-1 $\alpha$ , TIMP-1, TIMP-2, TNF $\beta$ , TGF $\beta$ 2, TGF $\beta$ 3, VEGF).

## 17. Species Reactivity

This kit recognizes human TGF beta 1.

Please contact our Technical Support team for more information.

## 18. Troubleshooting

| Problem             | Reason  | Solution  |
|---------------------|---|---|
| Poor standard curve | Inaccurate pipetting                            | Check pipettes  |
|                     | Improper standards dilution                     | Prior to opening, briefly spin the stock standard tube and dissolve the powder thoroughly by gentle mixing  |
| Low Signal          | Incubation times too brief                      | Ensure sufficient incubation times; change to overnight standard/sample incubation  |
|                     | Inadequate reagent volumes or improper dilution | Check pipettes and ensure correct preparation   |
| High background     | Plate is insufficiently washed                  | Review manual for proper wash technique. If using a plate washer, check all ports for obstructions  |
|                     | Contaminated wash buffer                        | Prepare fresh wash buffer   |
| Low sensitivity     | Improper storage of the ELISA kit               | Store the reconstituted protein at $-80^{\circ}\text{C}$ , all other assay components $4^{\circ}\text{C}$ . Keep substrate solution protected from light. |
|                     | Stop solution                                   | Stop solution should be added to each well before measure.  |

# 19. Notes







## Technical Support

Copyright © 2017 Abcam, All Rights Reserved. The Abcam logo is a registered trademark. All information / detail is correct at time of going to print.

### **Austria**

wissenschaftlicherdienst@abcam.com | 019-288-259

### **France**

supportscientifique@abcam.com | 01.46.94.62.96

### **Germany**

wissenschaftlicherdienst@abcam.com | 030-896-779-154

### **Spain**

soportecientifico@abcam.com | 91-114-65-60

### **Switzerland**

technical@abcam.com

Deutsch: 043-501-64-24 | Français: 061-500-05-30

### **UK, EU and ROW**

technical@abcam.com | +44(0)1223-696000

### **Canada**

ca.technical@abcam.com | 877-749-8807

### **US and Latin America**

us.technical@abcam.com | 888-772-2226

### **Asia Pacific**

hk.technical@abcam.com | (852) 2603-6823

### **China**

cn.technical@abcam.com | 400 921 0189 / +86 21 2070 0500

### **Japan**

technical@abcam.co.jp | +81-(0)3-6231-0940

### **Singapore**

sg.technical@abcam.com | 800 188-5244

### **Australia**

au.technical@abcam.com | +61-(0)3-8652-1450

### **New Zealand**

nz.technical@abcam.com | +64-(0)9-909-7829