

ab193736 – beta NGF Rat ELISA Kit

Instructions for Use

For the quantitative measurement of rat beta NGF cell and tissue lysates.

This product is for research use only and is not intended for diagnostic use.

Table of Contents

INTRODUCTION

1. BACKGROUND	2
2. ASSAY SUMMARY	3
3. PRECAUTIONS	4

GENERAL INFORMATION

4. STORAGE AND STABILITY	4
5. MATERIALS SUPPLIED	4
6. MATERIALS REQUIRED, NOT SUPPLIED	5
7. LIMITATIONS	5
8. TECHNICAL HINTS	5

ASSAY PREPARATION

9. REAGENT PREPARATION	7
10. STANDARD PREPARATIONS	8
11. SAMPLE PREPARATION	10
12. PLATE PREPARATION	10

ASSAY PROCEDURE

13. ASSAY PROCEDURE	11
---------------------	----

DATA ANALYSIS

14. CALCULATIONS	12
15. TYPICAL DATA	13
16. TYPICAL SAMPLE VALUES	14
17. ASSAY SPECIFICITY	15

RESOURCES

18. TROUBLESHOOTING	16
19. NOTES	17

1. BACKGROUND

Abcam's beta NGF Rat ELISA Kit (ab193736) is an *in vitro* enzyme-linked immunosorbent assay for the quantitative measurement of rat beta NGF in cell lysate and tissue lysate.

This assay employs an antibody specific for rat beta NGF coated on a 96-well plate. Standards and samples are pipetted into the wells and the immobilized antibody captures beta NGF present in the samples. The wells are washed and biotinylated anti-rat beta NGF antibody is added. After washing away any unbound biotinylated antibody, an HRP-conjugated streptavidin is pipetted to the wells. After incubation, the wells are again washed, followed by the addition of a TMB substrate solution to the wells. Color will develop in proportion to the amount of beta NGF bound in each well. Addition of the Stop Solution will change the color from blue to yellow, and the intensity of the color is measured at 450 nm.


2. ASSAY SUMMARY

Primary capture antibody



Prepare all reagents, samples and standards as instructed.

Sample



Add standard or sample to each well used. Incubate at room temperature

Primary detector antibody



Add prepared biotinylated antibody to each well. Incubate at room temperature.

Streptavidin Label



Add prepared streptavidin solution. Incubate at room temperature.

Substrate **Colored product**



Add TMB One-Step Development Solution to each well. Incubate at room temperature. Add Stop Solution to each well. Read immediately.

3. PRECAUTIONS

Please read these instructions carefully prior to beginning the assay.

Modifications to the kit components or procedures may result in loss of performance.

4. STORAGE AND STABILITY

Store kit at **-20°C** immediately upon receipt.

Refer to list of materials supplied for storage conditions of individual components. Observe the storage conditions for individual prepared components in sections 9 & 10.

5. MATERIALS SUPPLIED

Item	Amount	Storage Condition (Before Preparation)
Pre-coated beta NGF microplate (12 x 8 well strips)	96 wells	-20°C
20X Wash Buffer Concentrate	25 mL	-20°C
Rat beta NGF Standard (recombinant)	2 vials	-20°C
5X Assay Diluent	15 mL	-20°C
5X Sample Diluent Buffer	10 mL	-20°C
Detection Antibody beta NGF (biotinylated anti-rat beta NGF)	2 vials	-20°C
500X HRP-Streptavidin concentrate	200 µL	-20°C
TMB One-Step Substrate Reagent	12 mL	-20°C
Stop Solution: 0.2M sulfuric acid	8 mL	-20°C
2X Cell Lysate Buffer	5 mL	-20°C

6. MATERIALS REQUIRED, NOT SUPPLIED

These materials are not included in the kit, but will be required to successfully utilize this assay:

- Microplate reader capable of measuring absorbance at 450 nm.
- Precision pipettes to deliver 2 μ L to 1 mL volumes.
- Adjustable 1-25 mL pipettes for reagent preparation.
- 100 mL and 1 liter graduated cylinders.
- Absorbent paper.
- Distilled or deionized water.
- Log-log graph paper or computer and software for ELISA data analysis.
- Tubes to prepare standard or sample dilutions.

7. LIMITATIONS

- Do not mix or substitute reagents or materials from other kit lots or vendors.

8. TECHNICAL HINTS

- Samples which generate values that are greater than the most concentrated standard should be further diluted in the appropriate sample dilution buffer.
- Avoid foaming or bubbles when mixing or reconstituting components.
- Avoid cross contamination of samples or reagents by changing tips between sample, standard and reagent additions.
- Ensure plates are properly sealed or covered during incubation steps.
- Completely aspirate all solutions and buffers during wash steps. When preparing your standards, it is critical to briefly centrifuge the

vial first. The powder may adhere to the cap and not be included in the standard solution resulting in an incorrect concentration. Be sure to dissolve the powder thoroughly when reconstituting. After adding Assay Diluent to the vial, we recommend inverting the tube a few times, then flick the tube a few times, and centrifuge briefly; repeat this procedure 3-4 times. This is an effective technique for thorough mixing of the standard without using excessive mechanical force.

- Do not vortex the standard during reconstitution, as this will destabilize the protein.
- Once your standard has been reconstituted, it should be used right away or else frozen for later use.
- Keep the standard dilutions on ice during preparation, but the ELISA procedure should be done at room temperature.
- Be sure to discard the working standard dilutions after use – they do not store well.

This kit is sold based on number of tests. A ‘test’ simply refers to a single assay well. The number of wells that contain sample, control or standard will vary by product. Review the protocol completely to confirm this kit meets your requirements. Please contact our Scientific Support staff with any questions.

9. REAGENT PREPARATION

Equilibrate all reagents and samples to room temperature (18-25°C) prior to use.

9.1 1X Sample Diluent Buffer

Dilute 5X Sample Diluent Buffer 5-fold with deionized or distilled water before use.

9.2 1X Assay Diluent

Dilute 5X Assay Diluent 5-fold with deionized or distilled water before use.

9.3 1X Wash Buffer

If the 20X Wash Concentrate contains visible crystals, equilibrate to room temperature and mix gently until dissolved. Dilute 20 mL of 20X Wash Buffer Concentrate into 380 mL deionized or distilled water to yield 400 mL of 1X Wash Buffer.

9.4 Detection Antibody beta NGF (biotinylated anti-rat beta NGF)

Briefly centrifuge the Detection Antibody vial before use. Add 100 μ L of 1X Assay Diluent into the vial to prepare a detection antibody concentrate. Pipette up and down to mix gently (the concentrate can be stored at 4°C for 5 days). The detection antibody concentrate should be diluted 80-fold with 1X Assay Diluent B and used in Assay Procedure.

9.5 2X Cell Lysate Buffer

Dilute 2-fold with deionized or distilled water (for cell lysate and tissue lysate).

9.6 1X HRP-Streptavidin Solution

Briefly centrifuge the 500X HRP-Streptavidin concentrate vial and pipette up and down to mix gently before use. The 500X HRP-Streptavidin concentrate should be diluted 500-fold with 1X Assay Diluent.

For example: Briefly centrifuge the vial and pipette up and down to mix gently. Add 20 μ L of HRP-Streptavidin

concentrate into a tube with 10 mL 1X Assay Diluent B to prepare a 1X HRP-Streptavidin solution (do not store the diluted solution for next day use). Mix well.

10. STANDARD PREPARATIONS

- **Prepare serially diluted standards immediately prior to use. Always prepare a fresh set of standards for every use.**
- **Standard (recombinant protein) should be stored at -20°C or -80°C (recommended at -80°C) after reconstitution.**

10.1 Briefly centrifuge the vial of rat beta NGF Standard and then add 400 μ L 1X Sample Diluent Buffer into the rat beta NGF Standard vial to prepare a 50 ng/mL **Stock Standard**. Mix thoroughly but gently.

10.2 Label tubes #1-8.

10.3 Prepare the 15,000 pg/mL **Standard #1** by adding 150 μ L Stock Standard into tube #1 along with 350 μ L 1X Sample Diluent Buffer. Mix thoroughly but gently.

10.4 Add 400 μ L 1X Sample Diluent Buffer into tubes 2-8.

10.5 Prepare **Standard #2** by adding 200 μ L Standard #1 to tube #2. Mix thoroughly but gently.

10.6 Prepare **Standard #3** by adding 200 μ L from Standard #2 to tube #3. Mix thoroughly but gently.

10.7 Using the table below as a guide, prepare further serial dilutions.

10.8 Standard #8 contains no protein and is the Blank control.

ASSAY PREPARATION

Standard Dilution Preparation Table

Standard #	Sample to Dilute	Volume to Dilute (μL)	Volume of Diluent (μL)	Starting Conc. (pg/mL)	Final Conc. (pg/mL)
1	Stock	150	350	50,000	15,000
2	Standard #1	200	400	15,000	5,000
3	Standard #2	200	400	5,000	1,666
4	Standard #3	200	400	1,666	555.6
5	Standard #4	200	400	555.6	182.5
6	Standard #5	200	400	182.5	61.73
7	Standard #6	200	400	61.73	20.58
8 (Blank)	none	-	400	0	0



11. SAMPLE PREPARATION

- Sample Dilution: 1X Sample Diluent Buffer should be used for dilution of tissue and cell lysates.
- The recommended minimum dilution for tissue lysate and cell lysate is to achieve 1 mg/mL of protein or a 5-fold dilution.
- Please note that levels of the target protein may vary between different specimens. Optimal dilution factors for each sample must be determined by the investigator.

12. PLATE PREPARATION

- The 96 well plate strips included with this kit are supplied ready to use. It is not necessary to rinse the plate prior to adding reagents.
- Unused well strips should be returned to the plate packet and stored at 4°C.
- For each assay performed, a minimum of 2 wells must be used as blanks, omitting primary antibody from well additions.
- For statistical reasons, we recommend each sample should be assayed with a minimum of two replicates (duplicates).
- Well effects have not been observed with this assay.

13. ASSAY PROCEDURE

- **Equilibrate all materials and prepared reagents to room temperature (18 - 25°C) prior to use.**

It is recommended to assay all standards, controls and samples in duplicate.

- 13.1. Add 100 μL of each standard (see Standard Preparations, section) and sample into appropriate wells. Cover plate and incubate for 2.5 hours at room temperature or overnight at 4°C with gentle shaking.
- 13.2. Discard the solution and wash 4 times with 1X Wash Buffer. Wash by filling each well with 300 μL 1X Wash Buffer using a multi-channel pipette or automatic plate washer. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it by tapping gently against clean paper towels.
- 13.3. Add 100 μL of the prepared Detection Antibody beta NGF (see Reagent Preparation section) to each well. Incubate for 1 hour at room temperature with gentle shaking.
- 13.4. Discard the solution. Repeat the wash as in step 13.2.
- 13.5. Add 100 μL of prepared 1X HRP-Streptavidin solution (see Reagent Preparation section) to each well. Incubate for 45 minutes at room temperature with gentle shaking
- 13.6. Discard the solution. Repeat the wash as in step 13.2.
- 13.7. Add 100 μL of TMB One-Step Substrate Reagent to each well. Incubate for 30 minutes at room temperature in the dark with gentle shaking.
- 13.8. Add 50 μL of Stop Solution to each well. Read at 450 nm immediately.

14. CALCULATIONS

Calculate the mean absorbance for each set of duplicate standards, controls and samples, and subtract the average Blank absorbance value. Plot the standard curve on log-log graph paper, with standard concentration on the x-axis and absorbance on the y-axis. Draw the best-fit straight line through the standard points.

15. TYPICAL DATA

TYPICAL STANDARD CURVE – Data provided for **demonstration purposes only**. A new standard curve must be generated for each assay performed.




Figure 1. Example of typical rat beta NGF standard curve. The standard curve was prepared as described in Section 10.

16. TYPICAL SAMPLE VALUES

SENSITIVITY –

The minimum detectable dose of rat beta NGF is 15 pg/mL.

RECOVERY –

Recovery was determined by spiking various levels of rat beta NGF into tissue lysate and cell lysate. Mean recoveries are as follows:

Sample Type	Average % Recovery	Range (%)
Tissue lysate	99.34	70-121
Cell lysate	98.06	69-108

LINEARITY OF DILUTION -

Tissue lysate Dilution	Average % Expected Value	Range (%)
1:2	94.67	67-122
1:4	74.01	67-82

Cell lysate Dilution	Average % Expected Value	Range (%)
1:2	79.17	73-84
1:4	75.29	68-82

PRECISION –

	Intra-Assay	Inter-Assay
%CV	<10	<12

17. ASSAY SPECIFICITY

The antibodies used within this ELISA kit detect rat beta NGF.

Cross Reactivity: This ELISA kit shows no cross-reactivity with the following cytokines tested: rat CINC 2, CINC 3, CNTF, Fractalkine, IL-1 alpha, IL-1 beta, IL-4, IL-6, IL-10, GM-CSF, IFN gamma, Leptin, Lix, MCP 1, MIP 3 alpha, TIMP 1, TNF alpha.

Please contact our Scientific Support team for more information.

18. TROUBLESHOOTING

Problem	Cause	Solution
Poor standard curve	Inaccurate pipetting	Check pipette performance
	Improper standards dilution	Prior to opening, briefly spin the stock standard tube and dissolve the powder thoroughly by gentle mixing
Low Signal	Incubation times too brief	Ensure sufficient incubation time; change to overnight standard/sample incubation
	Inadequate reagent volumes or improper dilution	Check pipettes and ensure correct preparation
High %CV	Inaccurate pipetting	Check pipette performance
High background	Plate is insufficiently washed	Review manual for proper wash technique. If using a plate washer, ensure it is working properly.
	Contaminated wash buffer	Prepare fresh wash buffer
Low sensitivity	Improper storage of the ELISA kit	Store the reconstituted protein at -80°C, all other assay components 4°C. Keep substrate solution protected from light.
	Stop solution	Stop solution should be added to each well before measuring

19. NOTES

UK, EU and ROW

Email: technical@abcam.com | Tel: +44-(0)1223-696000

Austria

Email: wissenschaftlicherdienst@abcam.com | Tel: 019-288-259

France

Email: supportscientifique@abcam.com | Tel: 01-46-94-62-96

Germany

Email: wissenschaftlicherdienst@abcam.com | Tel: 030-896-779-154

Spain

Email: soportecientifico@abcam.com | Tel: 911-146-554

Switzerland

Email: technical@abcam.com

Tel (Deutsch): 0435-016-424 | Tel (Français): 0615-000-530

US and Latin America

Email: us.technical@abcam.com | Tel: 888-77-ABCAM (22226)

Canada

Email: ca.technical@abcam.com | Tel: 877-749-8807

China and Asia Pacific

Email: hk.technical@abcam.com | Tel: 400 921 0189 / +86 21 2070 0500

Japan

Email: technical@abcam.co.jp | Tel: +81-(0)3-6231-0940

www.abcam.com | www.abcam.cn | www.abcam.co.jp